



CERT

# DVGW CERT Conformity Certificate

## DVGW CERT Konformitätsbescheinigung

**AS-9501BM0119**  
Registration Number  
Registriernummer

<b>Field of Application</b> <i>Anwendungsbereich</i>	products for safe connection to the water supply <i>Anschlussssichere Produkte für das Trinkwassernetz</i>
<b>Manufacturer</b> <i>Hersteller</i>	AT-OS s.r.l. Viale del Lavoro 19, I-37030 Colognola ai Colli (VR)
<b>Distributor</b> <i>Vertreiber</i>	AT-OS s.r.l. Viale del Lavoro 19, I-37030 Colognola ai Colli (VR)
<b>Product Category</b> <i>Produktart</i>	devices connected to drinking water installation: products for medical use (0630)
<b>Product Description</b> <i>Produktbezeichnung</i>	bedpan washer
<b>Model</b> <i>Modell</i>	AF...
<b>Test Reports</b> <i>Prüfberichte</i>	laboratory control test: from 28.10.2015 (TZW) type testing: FA 010/11 + FA 011/10 from 13.12.2010 (TZW) type testing: FA 007/11 from 24.02.2011 (TZW)
<b>Test Basis</b> <i>Prüfgrundlagen</i>	DVGW W 540-VP (01.08.2010) BGA KTW (12.12.1985) DVGW W 270 (01.11.2007)

This product may be connected directly to the drinking water installation. The requirements for protection of drinking water against pollutions due to reflux are in compliance with DIN EN 1717.

**Date of Expiry / File No.** 23.04.2019 / 16-0182-WNV  
**Ablaufdatum / Aktenzeichen**

09.05.2016 GI A-1/2

Date, Issued by, Sheet, Head of Certification Body  
Datum, Bearbeiter, Blatt, Leiter der Zertifizierungsstelle

DVGW CERT GmbH is an accredited body by DAkkS according to DIN EN ISO/IEC 17065:2013 for certification of products for energy and water supply industry.

DVGW CERT GmbH ist von der DAkkS nach DIN EN ISO/IEC 17065:2013 akkreditierte Stelle für die Zertifizierung von Produkten der Energie- und Wasserversorgung.



DVGW CERT GmbH  
Zertifizierungsstelle

Josef-Wirmer-Str. 1-3  
53123 Bonn

Tel. +49 228 91 88 - 888  
Fax +49 228 91 88 - 993

www.dvgw-cert.com  
info@dvgw-cert.com

Type <i>Typ</i>	Technical Data <i>Technische Daten</i>	Remarks <i>Bemerkungen</i>
AF 2.60; AF 2.90, AF 60 FMS	type of protection unit: free outlet (type AB)	
AF 2.45; AF 2.45S	type of protection unit: free outlet (type AB)	
AF 2.75; AF 2.686	type of protection unit: free outlet (type AB)	

**certified Components*****zertifizierte Bauteile / Werkstoffe***

Registr. No. <i>Registrier.-Nr.</i>	Component <i>Bauteil (Produktart)</i>	Model/Type <i>Modell/Typ</i>	Manufacturer <i>Hersteller</i>
NW-6312AT2578	plug-in non-return valve	WZ 165-005; WZ 165-002; WZ 165-004 /WZ 165 - 004	A. u. K. Müller GmbH & Co KG
DW-0304BR0521	flexible pressure hose, class I, KTW-A	CPWX/CPWX 13	HYDROGOMMA S.P.A.
DW-0304CN0233	flexible pressure hose, class I, KTW-A	SPWX/SPWX	HYDROGOMMA S.P.A.

**Hints of Utilization / Remarks*****Verwendungshinweise / Bemerkungen***

The required protective device according to DIN EN 1717 is integral part of the tested component.

All materials in contact with drinking water are hygienic harmless up to the protection point.

This DVGW test certificate does not make any statements about the operation of the tested product.